

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023768.D
 Acq On : 17 Aug 2021 17:16
 Operator : JC/MD
 Sample : M3407-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 25N

Quant Time: Aug 17 17:41:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	144001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	242097	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90918	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	98877	49.197	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.400%
35) Dibromofluoromethane	5.397	113	79108	49.369	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.740%
50) Toluene-d8	8.653	98	304370	49.907	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.820%
62) 4-Bromofluorobenzene	11.085	95	101611	45.259	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.520%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	42729	47.549	ug/l	90
20) Methylene Chloride	2.794	84	13044	7.599	ug/l	96
25) 2-Butanone	4.580	43	10316	7.579	ug/l	89
43) Isopropyl Acetate	6.354	43	22725	5.272	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023768.D
Acq On : 17 Aug 2021 17:16
Operator : JC/MD
Sample : M3407-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
25N

Quant Time: Aug 17 17:41:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

