

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023771.D
 Acq On : 17 Aug 2021 18:27
 Operator : JC/MD
 Sample : M3389-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CRT-21

Quant Time: Aug 18 06:11:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	243923	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	225133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92678	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101446	51.466	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.940%
35) Dibromofluoromethane	5.397	113	78884	48.861	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.720%
50) Toluene-d8	8.653	98	298282	48.543	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.080%
62) 4-Bromofluorobenzene	11.085	95	102239	45.198	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.400%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	5750	6.524	ug/l #	83
20) Methylene Chloride	2.794	84	15401	9.148	ug/l	88
30) Chloroform	5.111	83	3464	1.182	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023771.D
Acq On : 17 Aug 2021 18:27
Operator : JC/MD
Sample : M3389-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CRT-21

Quant Time: Aug 18 06:11:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

