

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023774.D
 Acq On : 17 Aug 2021 19:38
 Operator : JC/MD
 Sample : PB138430TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138430TB

Quant Time: Aug 18 06:12:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	140569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	241542	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	237758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96091	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100757	51.357	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.720%
35) Dibromofluoromethane	5.397	113	78495	49.099	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.200%
50) Toluene-d8	8.653	98	309157	50.808	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.620%
62) 4-Bromofluorobenzene	11.085	95	106832	47.694	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.380%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	27522	31.374	ug/l	93
20) Methylene Chloride	2.794	84	13266	7.917	ug/l #	86
43) Isopropyl Acetate	6.354	43	20827	4.843	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023774.D
Acq On : 17 Aug 2021 19:38
Operator : JC/MD
Sample : PB138430TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138430TB

Quant Time: Aug 18 06:12:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

