

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023775.D
 Acq On : 17 Aug 2021 20:01
 Operator : JC/MD
 Sample : PB138478TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138478TB

Quant Time: Aug 18 06:12:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	140980	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	245499	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236140	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103030	52.362	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.720%	
35) Dibromofluoromethane	5.397	113	79771	49.093	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.180%	
50) Toluene-d8	8.653	98	312432	50.519	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.040%	
62) 4-Bromofluorobenzene	11.085	95	108052	47.461	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.920%	
Target Compounds						Qvalue
16) Acetone	2.386	43	27111	30.816	ug/l	99
20) Methylene Chloride	2.794	84	14489	8.622	ug/l	88
43) Isopropyl Acetate	6.355	43	21572	4.935	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023775.D
Acq On : 17 Aug 2021 20:01
Operator : JC/MD
Sample : PB138478TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138478TB

Quant Time: Aug 18 06:12:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

