

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023779.D
 Acq On : 17 Aug 2021 22:22
 Operator : JC/MD
 Sample : VX0817WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0817WBL02

Quant Time: Aug 18 06:14:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	142155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	245852	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	226783	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93893	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	99705	50.254	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.500%	
35) Dibromofluoromethane	5.397	113	80106	49.228	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.460%	
50) Toluene-d8	8.653	98	296945	47.946	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.900%	
62) 4-Bromofluorobenzene	11.085	95	104820	45.975	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.960%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023779.D
Acq On : 17 Aug 2021 22:22
Operator : JC/MD
Sample : VX0817WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0817WBL02

Quant Time: Aug 18 06:14:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

