

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023782.D
 Acq On : 17 Aug 2021 23:33
 Operator : JC/MD
 Sample : PB138478ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138478ZHE#01

Quant Time: Aug 18 07:00:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	152932	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	261499	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107971	50.585	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.180%		
35) Dibromofluoromethane	5.397	113	85398	49.340	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.680%		
50) Toluene-d8	8.653	98	330976	50.243	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 100.480%		
62) 4-Bromofluorobenzene	11.085	95	111080	45.805	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 91.620%		
Target Compounds						Qvalue
16) Acetone	2.386	43	29041	30.430	ug/l	93
20) Methylene Chloride	2.800	84	15230	8.354	ug/l	95
43) Isopropyl Acetate	6.355	43	22087	4.744	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023782.D
Acq On : 17 Aug 2021 23:33
Operator : JC/MD
Sample : PB138478ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138478ZHE#01

Quant Time: Aug 18 07:00:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

