

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023783.D
 Acq On : 17 Aug 2021 23:56
 Operator : JC/MD
 Sample : PB138478ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138478ZHE#02

Quant Time: Aug 18 07:01:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143093	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247350	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	101652	50.899	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	101.800%	
35) Dibromofluoromethane	5.397	113	80080	48.914	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.820%	
50) Toluene-d8	8.653	98	312050	50.080	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.160%	
62) 4-Bromofluorobenzene	11.085	95	107324	46.788	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	93.580%	
Target Compounds						Qvalue
16) Acetone	2.392	43	27713	31.035	ug/l	90
20) Methylene Chloride	2.800	84	14349	8.412	ug/l	96
43) Isopropyl Acetate	6.348	43	21538	4.890	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023783.D
Acq On : 17 Aug 2021 23:56
Operator : JC/MD
Sample : PB138478ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138478ZHE#02

Quant Time: Aug 18 07:01:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

