

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023784.D
 Acq On : 18 Aug 2021 00:20
 Operator : JC/MD
 Sample : M3398-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SWCS-05-0510

Quant Time: Aug 18 07:01:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	139644	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	241890	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232416	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102205	52.440	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.880%	
35) Dibromofluoromethane	5.397	113	78925	49.297	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.600%	
50) Toluene-d8	8.653	98	304102	49.906	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.820%	
62) 4-Bromofluorobenzene	11.085	95	104979	46.799	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.600%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	40063	45.973	ug/l	93
20) Methylene Chloride	2.794	84	13531	8.129	ug/l	91
25) 2-Butanone	4.580	43	8333	6.313	ug/l #	81
43) Isopropyl Acetate	6.354	43	19771	4.590	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023784.D
Acq On : 18 Aug 2021 00:20
Operator : JC/MD
Sample : M3398-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SWCS-05-0510

Quant Time: Aug 18 07:01:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

