

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023786.D
 Acq On : 18 Aug 2021 01:07
 Operator : JC/MD
 Sample : M3424-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18S

Quant Time: Aug 18 07:02:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	144044	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	248016	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	241348	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103192	51.329	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.660%	
35) Dibromofluoromethane	5.397	113	80601	49.100	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.200%	
50) Toluene-d8	8.653	98	315809	50.547	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.100%	
62) 4-Bromofluorobenzene	11.085	95	106373	46.249	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.500%	
Target Compounds						Qvalue
16) Acetone	2.392	43	13371	14.875	ug/l	90
20) Methylene Chloride	2.800	84	11367	6.620	ug/l	93
43) Isopropyl Acetate	6.354	43	20178	4.569	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023786.D
Acq On : 18 Aug 2021 01:07
Operator : JC/MD
Sample : M3424-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18S

Quant Time: Aug 18 07:02:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

