

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023787.D
 Acq On : 18 Aug 2021 01:30
 Operator : JC/MD
 Sample : M3424-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 19N

Quant Time: Aug 18 07:02:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	138040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	239191	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94782	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99874	51.839	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.680%	
35) Dibromofluoromethane	5.397	113	79457	50.189	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.380%	
50) Toluene-d8	8.653	98	301172	49.983	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.960%	
62) 4-Bromofluorobenzene	11.085	95	104090	46.926	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.860%	
Target Compounds						Qvalue
16) Acetone	2.392	43	12775	14.830	ug/l	92
20) Methylene Chloride	2.794	84	10811	6.570	ug/l	94
43) Isopropyl Acetate	6.348	43	19109	4.487	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023787.D
Acq On : 18 Aug 2021 01:30
Operator : JC/MD
Sample : M3424-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
19N

Quant Time: Aug 18 07:02:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

