

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023789.D
 Acq On : 18 Aug 2021 02:18
 Operator : JC/MD
 Sample : M3424-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20N

Quant Time: Aug 18 07:02:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	133847	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230526	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216148	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81748	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	94422	50.545	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.080%
35) Dibromofluoromethane	5.397	113	75383	49.406	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.820%
50) Toluene-d8	8.653	98	288435	49.668	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.340%
62) 4-Bromofluorobenzene	11.085	95	93679	43.820	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.640%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	12917	15.464	ug/l	89
20) Methylene Chloride	2.800	84	11316	7.093	ug/l #	96
43) Isopropyl Acetate	6.354	43	18034	4.393	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023789.D
Acq On : 18 Aug 2021 02:18
Operator : JC/MD
Sample : M3424-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20N

Quant Time: Aug 18 07:02:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

