

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023792.D
 Acq On : 18 Aug 2021 03:28
 Operator : JC/MD
 Sample : M3425-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 21S

Quant Time: Aug 18 07:03:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235954	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	99758	52.268	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.540%
35) Dibromofluoromethane	5.397	113	77461	49.600	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.200%
50) Toluene-d8	8.653	98	298416	50.205	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.400%
62) 4-Bromofluorobenzene	11.085	95	102928	47.039	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.080%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	12576	14.737	ug/l	94
20) Methylene Chloride	2.794	84	12235	7.506	ug/l	95
43) Isopropyl Acetate	6.348	43	18722	4.456	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023792.D
Acq On : 18 Aug 2021 03:28
Operator : JC/MD
Sample : M3425-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
21S

Quant Time: Aug 18 07:03:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

