

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081721\  
 Data File : VX023793.D  
 Acq On : 18 Aug 2021 03:52  
 Operator : JC/MD  
 Sample : M3425-12  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 44 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampled :  
 22N

Quant Time: Aug 18 07:04:02 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081721W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 17 11:57:51 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.562	168	138496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235800	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89303	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98149	50.776	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	101.560%	
35) Dibromofluoromethane	5.397	113	76886	49.264	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.520%	
50) Toluene-d8	8.653	98	301233	50.712	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	101.420%	
62) 4-Bromofluorobenzene	11.085	95	100352	45.892	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	91.780%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	13175	15.244	ug/l	95
20) Methylene Chloride	2.794	84	12498	7.570	ug/l	93
43) Isopropyl Acetate	6.355	43	18848	4.489	ug/l	92
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081721\  
Data File : VX023793.D  
Acq On : 18 Aug 2021 03:52  
Operator : JC/MD  
Sample : M3425-12  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 44 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
22N

Quant Time: Aug 18 07:04:02 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081721W.M  
Quant Title : SW846 8260  
QLast Update : Tue Aug 17 11:57:51 2021  
Response via : Initial Calibration

