

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023796.D
 Acq On : 18 Aug 2021 05:03
 Operator : JC/MD
 Sample : M3425-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 23S

Quant Time: Aug 18 07:04:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132646	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229050	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77442	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95638	51.659	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.320%	
35) Dibromofluoromethane	5.398	113	74115	48.888	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.780%	
50) Toluene-d8	8.653	98	289003	50.087	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.180%	
62) 4-Bromofluorobenzene	11.086	95	91653	43.149	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 86.300%	
Target Compounds						Qvalue
16) Acetone	2.386	43	13021	15.730	ug/l	89
20) Methylene Chloride	2.794	84	12095	7.649	ug/l	88
43) Isopropyl Acetate	6.355	43	17223	4.223	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023796.D
Acq On : 18 Aug 2021 05:03
Operator : JC/MD
Sample : M3425-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
23S

Quant Time: Aug 18 07:04:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

