

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023798.D
 Acq On : 18 Aug 2021 05:50
 Operator : JC/MD
 Sample : M3429-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-2

Quant Time: Aug 18 07:05:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230105	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	224053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89494	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95716	50.837	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.680%
35) Dibromofluoromethane	5.397	113	75086	49.301	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.600%
50) Toluene-d8	8.653	98	293973	50.714	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.420%
62) 4-Bromofluorobenzene	11.085	95	98642	46.226	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.460%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	13077	15.534	ug/l	91
20) Methylene Chloride	2.794	84	12260	7.624	ug/l #	87
43) Isopropyl Acetate	6.348	43	19175	4.680	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023798.D
Acq On : 18 Aug 2021 05:50
Operator : JC/MD
Sample : M3429-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-2

Quant Time: Aug 18 07:05:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

