

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023799.D
 Acq On : 18 Aug 2021 06:13
 Operator : JC/MD
 Sample : M3429-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-3

Quant Time: Aug 18 07:05:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	231848	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98176	52.320	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.640%
35) Dibromofluoromethane	5.397	113	75857	49.433	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.860%
50) Toluene-d8	8.653	98	296820	50.820	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.640%
62) 4-Bromofluorobenzene	11.085	95	100785	46.875	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.760%
Target Compounds						Qvalue
16) Acetone	2.386	43	14058	16.755	ug/l	91
20) Methylene Chloride	2.794	84	12253	7.645	ug/l	88
43) Isopropyl Acetate	6.355	43	19069	4.619	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023799.D
Acq On : 18 Aug 2021 06:13
Operator : JC/MD
Sample : M3429-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-3

Quant Time: Aug 18 07:05:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

