

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023801.D
 Acq On : 18 Aug 2021 07:00
 Operator : JC/MD
 Sample : M3441-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-5

Quant Time: Aug 18 07:48:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235450	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	226233	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97727	51.922	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.840%			
35) Dibromofluoromethane	5.397	113	76317	48.972	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.940%			
50) Toluene-d8	8.653	98	297073	50.086	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.180%			
62) 4-Bromofluorobenzene	11.085	95	98605	45.160	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 90.320%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	10158	12.070	ug/l	94
20) Methylene Chloride	2.794	84	8946	5.565	ug/l #	84
43) Isopropyl Acetate	6.354	43	18153	4.330	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023801.D
Acq On : 18 Aug 2021 07:00
Operator : JC/MD
Sample : M3441-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-5

Quant Time: Aug 18 07:48:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

