

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023778.D
 Acq On : 17 Aug 2021 21:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 18 06:13:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.528	0.544	-3.0	86	0.00
3 P	Chloromethane	0.579	0.583	-0.7	86	0.00
4 C	Vinyl Chloride	0.565	0.569	-0.7#	84	0.00
5 T	Bromomethane	0.355	0.364	-2.5	91	0.00
6 T	Chloroethane	0.383	0.371	3.1	87	0.00
7 T	Trichlorofluoromethane	0.892	0.928	-4.0	88	0.00
8 T	Diethyl Ether	0.311	0.324	-4.2	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.502	0.508	-1.2	85	0.00
10 T	Methyl Iodide	0.699	0.727	-4.0	84	0.00
11 T	Tert butyl alcohol	0.141	0.149	-5.7	89	0.00
12 CM	1,1-Dichloroethene	0.480	0.504	-5.0#	89	0.00
13 T	Acrolein	0.084	0.079	6.0	81	0.00
14 T	Allyl chloride	0.884	0.896	-1.4	86	0.00
15 T	Acrylonitrile	0.314	0.328	-4.5	88	0.00
16 T	Acetone	0.312	0.301	3.5	85	0.00
17 T	Carbon Disulfide	1.144	1.173	-2.5	84	0.00
18 T	Methyl Acetate	0.779	0.807	-3.6	90	0.00
19 T	Methyl tert-butyl Ether	1.670	1.730	-3.6	86	0.00
20 T	Methylene Chloride	0.596	0.581	2.5	88	0.00
21 T	trans-1,2-Dichloroethene	0.513	0.533	-3.9	88	0.00
22 T	Diisopropyl ether	1.761	1.816	-3.1	86	0.00
23 T	Vinyl Acetate	1.425	1.502	-5.4	86	0.00
24 P	1,1-Dichloroethane	0.980	1.004	-2.4	87	0.00
25 T	2-Butanone	0.473	0.491	-3.8	87	0.00
26 T	2,2-Dichloropropane	0.805	0.708	12.0	74	0.00
27 T	cis-1,2-Dichloroethene	0.610	0.613	-0.5	85	0.00
28 T	Bromochloromethane	0.446	0.436	2.2	81	0.00
29 T	Tetrahydrofuran	0.306	0.320	-4.6	87	0.00
30 C	Chloroform	1.038	1.038	0.0	86	0.00
31 T	Cyclohexane	0.897	0.913	-1.8	86	0.00
32 T	1,1,1-Trichloroethane	0.889	0.917	-3.1	85	0.00
33 S	1,2-Dichloroethane-d4	0.698	0.691	1.0	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.331	0.326	1.5	86	0.00
36 T	1,1-Dichloropropene	0.463	0.480	-3.7	86	0.00
37 T	Ethyl Acetate	0.569	0.574	-0.9	87	0.00
38 T	Carbon Tetrachloride	0.483	0.495	-2.5	85	0.00
39 T	Methylcyclohexane	0.564	0.589	-4.4	85	0.00
40 TM	Benzene	1.369	1.402	-2.4	86	0.00
41 T	Methacrylonitrile	0.309	0.317	-2.6	87	0.00
42 TM	1,2-Dichloroethane	0.538	0.553	-2.8	86	0.00
43 T	Isopropyl Acetate	0.890	0.909	-2.1	86	0.00
44 TM	Trichloroethene	0.378	0.378	0.0	86	0.00
45 C	1,2-Dichloropropane	0.364	0.365	-0.3#	84	0.00
46 T	Dibromomethane	0.253	0.258	-2.0	86	0.00
47 T	Bromodichloromethane	0.458	0.476	-3.9	84	0.00
48 T	Methyl methacrylate	0.434	0.459	-5.8	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
 Data File : VX023778.D
 Acq On : 17 Aug 2021 21:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 18 06:13:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	90	0.00
50 S	Toluene-d8	1.260	1.240	1.6	85	0.00
51 T	4-Methyl-2-Pentanone	0.576	0.624	-8.3	88	0.00
52 CM	Toluene	0.887	0.907	-2.3#	85	0.00
53 T	t-1,3-Dichloropropene	0.521	0.540	-3.6	83	0.00
54 T	cis-1,3-Dichloropropene	0.555	0.564	-1.6	83	0.00
55 T	1,1,2-Trichloroethane	0.365	0.380	-4.1	88	0.00
56 T	Ethyl methacrylate	0.569	0.610	-7.2	86	0.00
57 T	1,3-Dichloropropane	0.606	0.623	-2.8	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.294	0.297	-1.0	85	0.00
59 T	2-Hexanone	0.442	0.491	-11.1	88	0.00
60 T	Dibromochloromethane	0.353	0.378	-7.1	83	0.00
61 T	1,2-Dibromoethane	0.386	0.402	-4.1	85	0.00
62 S	4-Bromofluorobenzene	0.464	0.479	-3.2	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.399	0.394	1.3	87	0.00
65 PM	Chlorobenzene	1.043	1.044	-0.1	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.377	-1.6	85	0.00
67 C	Ethyl Benzene	1.822	1.845	-1.3#	85	0.00
68 T	m/p-Xylenes	0.694	0.715	-3.0	86	0.00
69 T	o-Xylene	0.682	0.709	-4.0	87	0.00
70 T	Styrene	1.111	1.172	-5.5	86	0.00
71 P	Bromoform	0.263	0.280	-6.5	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.814	3.963	-3.9	87	0.00
74 T	N-amyl acetate	1.724	1.801	-4.5	86	0.00
75 P	1,1,2,2-Tetrachloroethane	1.250	1.247	0.2	88	0.00
76 T	1,2,3-Trichloropropane	1.166	1.157	0.8	85	0.00
77 T	Bromobenzene	0.939	0.969	-3.2	87	0.00
78 T	n-propylbenzene	4.362	4.556	-4.4	86	0.00
79 T	2-Chlorotoluene	2.636	2.727	-3.5	86	0.00
80 T	1,3,5-Trimethylbenzene	3.134	3.329	-6.2	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.385	3.0	80	0.00
82 T	4-Chlorotoluene	2.993	3.118	-4.2	85	0.00
83 T	tert-Butylbenzene	3.044	3.191	-4.8	86	0.00
84 T	1,2,4-Trimethylbenzene	3.125	3.279	-4.9	86	0.00
85 T	sec-Butylbenzene	3.854	4.131	-7.2	86	0.00
86 T	p-Isopropyltoluene	3.172	3.395	-7.0	85	0.00
87 T	1,3-Dichlorobenzene	1.699	1.751	-3.1	87	0.00
88 T	1,4-Dichlorobenzene	1.695	1.735	-2.4	86	0.00
89 T	n-Butylbenzene	2.832	2.959	-4.5	84	0.00
90 T	Hexachloroethane	0.488	0.547	-12.1	85	0.00
91 T	1,2-Dichlorobenzene	1.646	1.685	-2.4	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.278	-8.2	87	0.00
93 T	1,2,4-Trichlorobenzene	0.991	1.014	-2.3	85	0.00
94 T	Hexachlorobutadiene	0.442	0.439	0.7	86	0.00
95 T	Naphthalene	3.231	3.595	-11.3	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081721\
Data File : VX023778.D
Acq On : 17 Aug 2021 21:35
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 18 06:13:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.987	1.043	-5.7	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5