

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081722\
 Data File : VX030669.D
 Acq On : 17 Aug 2022 18:45
 Operator : JC/MD
 Sample : N4221-01ME
 Misc : 5.01g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0KP2ME

Quant Time: Aug 18 04:25:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 18 04:21:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	181707	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	163684	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	69407	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	87478	74.967	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	149.940%#
7) Chloroethane-d5	1.654	69	24535	29.826	ug/L	-0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	59.660%#
11) 1,1-Dichloroethene-d2	2.276	63	126325	57.231	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	114.460%
21) 2-Butanone-d5	4.501	46	117808	108.962	ug/L	0.04
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.960%
24) Chloroform-d	5.062	84	151884	63.230	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	126.460%#
26) 1,2-Dichloroethane-d4	5.958	65	104403	64.011	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	128.020%#
32) Benzene-d6	5.971	84	337662	72.017	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	144.040%#
36) 1,2-Dichloropropane-d6	7.312	67	105909	69.621	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	139.240%#
41) Toluene-d8	8.653	98	297012	67.203	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	134.400%#
43) trans-1,3-Dichloroprop...	8.952	79	42757	56.499	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	113.000%
47) 2-Hexanone-d5	9.403	63	83606	107.824	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.820%
56) 1,1,2,2-Tetrachloroeth...	11.201	84	127759	63.931	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	127.860%#
66) 1,2-Dichlorobenzene-d4	12.323	152	101893	75.688	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	151.380%#
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.294	96	3451	3.393	ug/L #	1
34) Trichloroethene	7.123	95	34039	22.394	ug/L	91
46) Tetrachloroethene	9.275	164	20465	21.304	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081722\
Data File : VX030669.D
Acq On : 17 Aug 2022 18:45
Operator : JC/MD
Sample : N4221-01ME
Misc : 5.01g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0KP2ME

Quant Time: Aug 18 04:25:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 18 04:21:09 2022
Response via : Initial Calibration

