

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081723\
 Data File : VX037060.D
 Acq On : 17 Aug 2023 14:46
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: Aug 18 06:05:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 05:51:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	171914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	297852	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268631	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	128371	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	41535	16.758	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	33.520%#
35) Dibromofluoromethane	5.385	113	32559	16.506	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	33.020%#
50) Toluene-d8	8.647	98	120141	16.790	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	33.580%#
62) 4-Bromofluorobenzene	11.079	95	46288	16.461	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	32.920%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	46267	23.168	ug/l	97
3) Chloromethane	1.301	50	36052	20.636	ug/l	94
4) Vinyl Chloride	1.374	62	40317	21.201	ug/l	99
5) Bromomethane	1.599	94	21984	23.208	ug/l	99
6) Chloroethane	1.679	64	25190	22.510	ug/l	100
7) Trichlorofluoromethane	1.886	101	70096	22.502	ug/l	99
8) Diethyl Ether	2.136	74	24341	20.197	ug/l	74
9) 1,1,2-Trichlorotrifluo...	2.325	101	38596	21.238	ug/l	96
10) Methyl Iodide	2.453	142	45556	20.626	ug/l	96
11) Tert butyl alcohol	2.965	59	59260	101.929	ug/l #	93
12) 1,1-Dichloroethene	2.319	96	37215	21.029	ug/l	85
13) Acrolein	2.240	56	41846	94.613	ug/l	99
14) Allyl chloride	2.666	41	57647	21.046	ug/l #	94
15) Acrylonitrile	3.063	53	106424	102.557	ug/l	99
16) Acetone	2.380	43	99306	102.420	ug/l	92
17) Carbon Disulfide	2.508	76	91589	20.347	ug/l	98
18) Methyl Acetate	2.703	43	75102	20.572	ug/l #	89
19) Methyl tert-butyl Ether	3.117	73	138341	20.612	ug/l	98
20) Methylene Chloride	2.788	84	41084	18.948	ug/l	87
21) trans-1,2-Dichloroethene	3.093	96	41546	20.652	ug/l	88
22) Diisopropyl ether	3.764	45	116269	20.685	ug/l #	87
23) Vinyl Acetate	3.721	43	453075	105.187	ug/l #	92
24) 1,1-Dichloroethane	3.611	63	73192	20.765	ug/l	97
25) 2-Butanone	4.556	43	150845	103.324	ug/l #	88
26) 2,2-Dichloropropane	4.477	77	65987	20.962	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	49243	20.763	ug/l	90
28) Bromochloromethane	4.898	49	31091	19.745	ug/l #	78
29) Tetrahydrofuran	5.007	42	96065	103.181	ug/l #	84
30) Chloroform	5.093	83	81049	20.484	ug/l	99
31) Cyclohexane	5.471	56	59867	21.482	ug/l #	81
32) 1,1,1-Trichloroethane	5.385	97	74485	21.007	ug/l	97
36) 1,1-Dichloropropene	5.690	75	58023	20.704	ug/l	98
37) Ethyl Acetate	4.715	43	57485	20.080	ug/l #	93
38) Carbon Tetrachloride	5.678	117	63769	20.947	ug/l	99
39) Methylcyclohexane	7.385	83	62985	21.391	ug/l	89
40) Benzene	6.038	78	167610	20.853	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081723\
 Data File : VX037060.D
 Acq On : 17 Aug 2023 14:46
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: Aug 18 06:05:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 05:51:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	30907	20.768	ug/l #	88
42) 1,2-Dichloroethane	6.086	62	65165	20.674	ug/l	96
43) Isopropyl Acetate	6.342	43	94573	20.808	ug/l #	92
44) Trichloroethene	7.123	130	46815	20.541	ug/l	93
45) 1,2-Dichloropropane	7.434	63	41343	20.861	ug/l	97
46) Dibromomethane	7.586	93	31558	20.213	ug/l	93
47) Bromodichloromethane	7.824	83	60945	20.380	ug/l	96
48) Methyl methacrylate	7.696	41	45672	20.605	ug/l #	87
49) 1,4-Dioxane	7.659	88	24136	404.322	ug/l #	86
51) 4-Methyl-2-Pentanone	8.574	43	293917	106.496	ug/l	93
52) Toluene	8.720	92	110221	20.755	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	65694	20.632	ug/l	96
54) cis-1,3-Dichloropropene	8.366	75	70775	20.569	ug/l	90
55) 1,1,2-Trichloroethane	9.153	97	44695	20.668	ug/l	97
56) Ethyl methacrylate	9.116	69	70106	20.697	ug/l	89
57) 1,3-Dichloropropane	9.311	76	74694	20.490	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.245	63	165000	101.507	ug/l	93
59) 2-Hexanone	9.433	43	222005	105.499	ug/l	91
60) Dibromochloromethane	9.519	129	47069	20.444	ug/l	99
61) 1,2-Dibromoethane	9.610	107	49371	20.852	ug/l	98
64) Tetrachloroethene	9.275	164	40987	20.915	ug/l	97
65) Chlorobenzene	10.080	112	118145	20.809	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	44406	21.191	ug/l	99
67) Ethyl Benzene	10.195	91	208637	20.884	ug/l	99
68) m/p-Xylenes	10.299	106	159797	41.419	ug/l	100
69) o-Xylene	10.640	106	79097	20.780	ug/l	100
70) Styrene	10.653	104	130314	20.784	ug/l	98
71) Bromoform	10.799	173	32642	20.825	ug/l #	99
73) Isopropylbenzene	10.964	105	203612	21.822	ug/l	98
74) N-amyl acetate	10.842	43	78818	21.740	ug/l	92
75) 1,1,2,2-Tetrachloroethane	11.214	83	65573	20.868	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	75412	27.584	ug/l	90
77) Bromobenzene	11.195	156	48024	21.074	ug/l	94
78) n-propylbenzene	11.305	91	227389	21.886	ug/l	100
79) 2-Chlorotoluene	11.366	91	141475	21.308	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	170847	21.950	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	19502	19.924	ug/l	90
82) 4-Chlorotoluene	11.451	91	163264	21.302	ug/l	98
83) tert-Butylbenzene	11.713	119	159549	21.494	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	168728	21.449	ug/l	97
85) sec-Butylbenzene	11.890	105	190191	21.936	ug/l	98
86) p-Isopropyltoluene	12.006	119	162994	21.838	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	88362	20.711	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	89964	20.758	ug/l	99
89) n-Butylbenzene	12.329	91	132866	21.471	ug/l	98
90) Hexachloroethane	12.536	117	26382	21.177	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	86096	20.608	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.945	75	14774	20.234	ug/l	82
93) 1,2,4-Trichlorobenzene	13.585	180	49623	20.300	ug/l	96
94) Hexachlorobutadiene	13.725	225	19293	20.430	ug/l	97
95) Naphthalene	13.774	128	203523	19.007	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	49489	20.377	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081723\
Data File : VX037060.D
Acq On : 17 Aug 2023 14:46
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Quant Time: Aug 18 06:05:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 05:51:31 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:168 Resp: 171914
Ion Ratio Lower Upper
168 100
99 64.5 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 23.168 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 85 Resp: 46267
Ion Ratio Lower Upper
85 100
87 32.0 16.7 50.1









#3
Chloromethane
Concen: 20.636 ug/l
RT: 1.301 min Scan# 35
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 50 Resp: 36052
Ion Ratio Lower Upper
50 100
52 35.4 25.5 38.3



#4
Vinyl Chloride
Concen: 21.201 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 62 Resp: 40317
Ion Ratio Lower Upper
62 100
64 32.1 25.4 38.0









#5
Bromomethane
Concen: 23.208 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 94 Resp: 21984
Ion Ratio Lower Upper
94 100
96 94.6 74.6 111.8



#6
Chloroethane
Concen: 22.510 ug/l
RT: 1.679 min Scan# 97
Delta R.T. 0.007 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 64 Resp: 25190
Ion Ratio Lower Upper
64 100
66 31.6 25.1 37.7









#7
 Trichlorofluoromethane
 Concen: 22.502 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020


Tgt Ion:101 Resp: 70096
 Ion Ratio Lower Upper
 101 100
 103 63.9 51.4 77.2



#8
 Diethyl Ether
 Concen: 20.197 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46

Tgt Ion: 74 Resp: 24341
 Ion Ratio Lower Upper
 74 100
 45 79.9 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.238 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:101 Resp: 38596
Ion Ratio Lower Upper
101 100
85 44.4 37.0 55.6
151 76.9 58.6 87.8



#10
Methyl Iodide
Concen: 20.626 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.006 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion:142 Resp: 45556
Ion Ratio Lower Upper
142 100
127 48.5 41.6 62.4
141 14.7 11.6 17.4









#11
Tert butyl alcohol
Concen: 101.929 ug/l
RT: 2.965 min Scan# 30
Delta R.T. -0.012 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 59 Resp: 59260
Ion Ratio Lower Upper
59 100
57 9.9 6.0 9.0#



#12
1,1-Dichloroethene
Concen: 21.029 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.006 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 96 Resp: 37215
Ion Ratio Lower Upper
96 100
61 154.4 144.6 216.8
98 65.4 50.4 75.6





#13

Acrolein

Concen: 94.613 ug/l

RT: 2.240 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 56 Resp: 41846

Ion Ratio Lower Upper

56 100

55 70.4 57.1 85.7



#14

Allyl chloride

Concen: 21.046 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.006 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46




Tgt Ion: 41 Resp: 57647


Ion Ratio Lower Upper

41 100

39 70.5 55.8 83.6

76 39.8 24.3 36.5#









#15
Acrylonitrile
Concen: 102.557 ug/l
RT: 3.063 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 53 Resp: 106424
Ion Ratio Lower Upper
53 100
52 83.2 66.0 99.0
51 37.5 29.8 44.6



#16
Acetone
Concen: 102.420 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 43 Resp: 99306
Ion Ratio Lower Upper
43 100
58 31.7 21.9 32.9









#17
Carbon Disulfide
Concen: 20.347 ug/l
RT: 2.508 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 76 Resp: 91589
Ion Ratio Lower Upper
76 100
78 9.3 7.0 10.6



#18
Methyl Acetate
Concen: 20.572 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 43 Resp: 75102
Ion Ratio Lower Upper
43 100
74 27.5 17.8 26.8#









#19
Methyl tert-butyl Ether
Concen: 20.612 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 73 Resp: 138341
Ion Ratio Lower Upper
73 100
57 20.5 17.3 25.9



#20
Methylene Chloride
Concen: 18.948 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 84 Resp: 41084
Ion Ratio Lower Upper
84 100
49 108.0 103.6 155.4
51 33.9 33.8 50.6
86 63.5 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 20.652 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 96 Resp: 41546
Ion Ratio Lower Upper
96 100
61 128.9 119.2 178.8
98 63.4 48.7 73.1



#22
Diisopropyl ether
Concen: 20.685 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 45 Resp: 116269
Ion Ratio Lower Upper
45 100
43 74.3 69.7 104.5
87 33.0 20.8 31.2#
59 12.1 9.0 13.4









#23
 Vinyl Acetate
 Concen: 105.187 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion: 43 Resp: 453075
 Ion Ratio Lower Upper
 43 100
 86 12.7 7.8 11.8#



#24
 1,1-Dichloroethane
 Concen: 20.765 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.006 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46

Tgt Ion: 63 Resp: 73192
 Ion Ratio Lower Upper
 63 100
 98 7.6 3.3 9.8
 100 5.1 2.1 6.5





#25

2-Butanone

Concen: 103.324 ug/l

RT: 4.556 min Scan# 5

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 43 Resp: 150845

Ion Ratio Lower Upper

43 100

72 30.3 19.4 29.2#



#26

2,2-Dichloropropane

Concen: 20.962 ug/l

RT: 4.477 min Scan# 556

Delta R.T. 0.000 min

Lab File: VX037060.D




Acq: 17 Aug 2023 14:46


Tgt Ion: 77 Resp: 65987

Ion Ratio Lower Upper

77 100

97 22.3 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 20.763 ug/l
 RT: 4.489 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion:	96	Resp:	49243
Ion	Ratio	Lower	Upper
96	100		
61	137.4	0.0	312.2
98	64.5	0.0	129.0



#28
 Bromochloromethane
 Concen: 19.745 ug/l
 RT: 4.898 min Scan# 625
 Delta R.T. -0.006 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46

Tgt Ion:	49	Resp:	31091
Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	3.8
130	91.1	57.5	86.3





#29

Tetrahydrofuran

Concen: 103.181 ug/l

RT: 5.007 min Scan# 643

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 42 Resp: 96065

Ion Ratio Lower Upper

42 100

72 53.1 34.2 51.4#

71 50.1 32.2 48.2#



#30

Chloroform

Concen: 20.484 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.000 min

Lab File: VX037060.D




Acq: 17 Aug 2023 14:46


Tgt Ion: 83 Resp: 81049

Ion Ratio Lower Upper

83 100

85 63.6 50.3 75.5









#31
Cyclohexane
Concen: 21.482 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 56 Resp: 59867
Ion Ratio Lower Upper
56 100
69 35.9 23.4 35.2#
84 101.0 65.9 98.9#



#32
1,1,1-Trichloroethane
Concen: 21.007 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 97 Resp: 74485
Ion Ratio Lower Upper
97 100
99 64.4 51.1 76.7
61 41.6 36.4 54.6









#33
 1,2-Dichloroethane-d4
 Concen: 16.758 ug/l
 RT: 5.958 min Scan# 799
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion: 65 Resp: 41535
 Ion Ratio Lower Upper
 65 100
 67 51.3 0.0 100.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.763 min Scan# 931
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46

Tgt Ion: 114 Resp: 297852
 Ion Ratio Lower Upper
 114 100
 63 19.0 0.0 45.0
 88 16.4 0.0 36.6









#35
Dibromofluoromethane
Concen: 16.506 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:113 Resp: 32559
Ion Ratio Lower Upper
113 100
111 103.8 82.9 124.3
192 18.3 13.3 19.9



#36
1,1-Dichloropropene
Concen: 20.704 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.006 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 75 Resp: 58023
Ion Ratio Lower Upper
75 100
110 36.2 17.0 50.9
77 30.8 24.7 37.1





#37

Ethyl Acetate

Concen: 20.080 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 43 Resp: 57485

Ion Ratio Lower Upper

43 100

61 15.2 10.6 16.0

70 13.7 8.3 12.5#



#38

Carbon Tetrachloride

Concen: 20.947 ug/l

RT: 5.678 min Scan# 753

Delta R.T. 0.006 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46




Tgt Ion: 117 Resp: 63769


Ion Ratio Lower Upper

117 100

119 97.1 78.2 117.2

121 32.3 24.4 36.6





#39

Methylcyclohexane

Concen: 21.391 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.006 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 83 Resp: 62985

Ion Ratio Lower Upper

83 100

55 68.1 64.6 97.0

98 48.5 36.7 55.1



#40

Benzene

Concen: 20.853 ug/l

RT: 6.038 min Scan# 812

Delta R.T. 0.000 min

Lab File: VX037060.D




Acq: 17 Aug 2023 14:46


Tgt Ion: 78 Resp: 167610

Ion Ratio Lower Upper

78 100

77 23.9 19.3 28.9









#41
Methacrylonitrile
Concen: 20.768 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion:	41	Resp:	30907
Ion	Ratio	Lower	Upper
41	100		
39	54.3	45.2	67.8
67	82.4	52.6	78.8#
52	33.0	22.7	34.1



#42
1,2-Dichloroethane
Concen: 20.674 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion:	62	Resp:	65165
Ion	Ratio	Lower	Upper
62	100		
98	9.3	0.0	16.0









#43
Isopropyl Acetate
Concen: 20.808 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 94573
Ion Ratio Lower Upper
43 100
61 21.8 15.2 22.8
87 16.1 9.8 14.6#



#44
Trichloroethene
Concen: 20.541 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion:130 Resp: 46815
Ion Ratio Lower Upper
130 100
95 95.0 0.0 204.4









#45
1,2-Dichloropropane
Concen: 20.861 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 63 Resp: 41343
Ion Ratio Lower Upper
63 100
65 32.1 24.3 36.5



#46
Dibromomethane
Concen: 20.213 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 93 Resp: 31558
Ion Ratio Lower Upper
93 100
95 84.3 66.6 100.0
174 98.2 69.4 104.0









#47
Bromodichloromethane
Concen: 20.380 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 60945
Ion Ratio Lower Upper
83 100
85 65.8 49.8 74.6
127 8.7 7.1 10.7



#48
Methyl methacrylate
Concen: 20.605 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 41 Resp: 45672
Ion Ratio Lower Upper
41 100
69 91.8 59.2 88.8#
39 61.6 47.7 71.5





#49

1,4-Dioxane

Concen: 404.322 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 88 Resp: 24136

Ion Ratio Lower Upper

88 100

43 29.6 0.0 0.0#

58 63.3 60.1 90.1



#50

Toluene-d8

Concen: 16.790 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX037060.D




Acq: 17 Aug 2023 14:46


Tgt Ion: 98 Resp: 120141

Ion Ratio Lower Upper

98 100

100 67.2 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 106.496 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 293917
Ion Ratio Lower Upper
43 100
58 42.3 30.4 45.6



#52
Toluene
Concen: 20.755 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 92 Resp: 110221
Ion Ratio Lower Upper
92 100
91 169.1 134.2 201.2









#53
 t-1,3-Dichloropropene
 Concen: 20.632 ug/l
 RT: 8.976 min Scan# 1294
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion: 75 Resp: 65694
 Ion Ratio Lower Upper
 75 100
 77 30.0 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 20.569 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46

Tgt Ion: 75 Resp: 70775
 Ion Ratio Lower Upper
 75 100
 77 33.5 25.3 37.9
 39 47.7 46.7 70.1





#55

1,1,2-Trichloroethane

Concen: 20.668 ug/l

RT: 9.153 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 97 Resp: 44695





Ion Ratio Lower Upper

97 100

83 81.8 65.8 98.8

85 56.0 41.6 62.4

99 63.2 47.8 71.8



#56

Ethyl methacrylate

Concen: 20.697 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46




Tgt Ion: 69 Resp: 70106


Ion Ratio Lower Upper

69 100

41 60.0 56.8 85.2

39 35.6 32.0 48.0





#57

1,3-Dichloropropane

Concen: 20.490 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.006 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 76 Resp: 74694

Ion Ratio Lower Upper

76 100

78 33.1 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 101.507 ug/l

RT: 8.245 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX037060.D




Acq: 17 Aug 2023 14:46


Tgt Ion: 63 Resp: 165000

Ion Ratio Lower Upper

63 100

106 31.7 22.5 33.7









#59
2-Hexanone
Concen: 105.499 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 222005
Ion Ratio Lower Upper
43 100
58 58.9 26.2 78.5



#60
Dibromochloromethane
Concen: 20.444 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 129 Resp: 47069
Ion Ratio Lower Upper
129 100
127 78.7 38.8 116.4









#61
1,2-Dibromoethane
Concen: 20.852 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion:107 Resp: 49371
Ion Ratio Lower Upper
107 100
109 92.5 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 16.461 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 95 Resp: 46288
Ion Ratio Lower Upper
95 100
174 72.0 0.0 135.6
176 69.8 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion:117 Resp: 268631
Ion Ratio Lower Upper
117 100
82 56.2 46.2 69.4
119 32.6 25.5 38.3



#64
Tetrachloroethene
Concen: 20.915 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion:164 Resp: 40987
Ion Ratio Lower Upper
164 100
166 133.5 102.7 154.1
129 99.7 80.4 120.6
131 97.4 76.5 114.7





#65

Chlorobenzene

Concen: 20.809 ug/l

RT: 10.080 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:112 Resp: 118145

Ion Ratio Lower Upper

112 100

114 31.4 25.7 38.5



#66

1,1,1,2-Tetrachloroethane

Concen: 21.191 ug/l

RT: 10.165 min Scan# 1489

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46




Tgt Ion:131 Resp: 44406


Ion Ratio Lower Upper

131 100

133 95.3 47.1 141.4

119 63.4 32.2 96.6









#67
Ethyl Benzene
Concen: 20.884 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 91 Resp: 208637
Ion Ratio Lower Upper
91 100
106 31.5 24.9 37.3



#68
m/p-Xylenes
Concen: 41.419 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 106 Resp: 159797
Ion Ratio Lower Upper
106 100
91 206.6 165.5 248.3





#69

o-Xylene

Concen: 20.780 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:106 Resp: 79097

Ion Ratio Lower Upper

106 100

91 218.5 108.9 326.8



#70

Styrene

Concen: 20.784 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.001 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46




Tgt Ion:104 Resp: 130314


Ion Ratio Lower Upper

104 100

78 52.1 43.6 65.4

103 54.3 43.6 65.4









#71
Bromoform
Concen: 20.825 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion:173 Resp: 32642
Ion Ratio Lower Upper
173 100
175 47.6 24.1 72.4
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.006 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion:152 Resp: 128371
Ion Ratio Lower Upper
152 100
115 73.3 43.3 129.9
150 163.4 0.0 345.0









#73
Isopropylbenzene
Concen: 21.822 ug/l
RT: 10.964 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:105 Resp: 203612
Ion Ratio Lower Upper
105 100
120 26.6 12.9 38.6



#74
N-ethyl acetate
Concen: 21.740 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion: 43 Resp: 78818
Ion Ratio Lower Upper
43 100
70 48.7 32.9 49.3
55 25.5 19.3 28.9
61 26.5 18.8 28.2









#75
 1,1,2,2-Tetrachloroethane
 Concen: 20.868 ug/l
 RT: 11.214 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion: 83 Resp: 65573
 Ion Ratio Lower Upper
 83 100
 131 10.7 5.3 15.8
 85 64.4 32.0 96.2



#76
 1,2,3-Trichloropropane
 Concen: 27.584 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46

Tgt Ion: 75 Resp: 75412
 Ion Ratio Lower Upper
 75 100
 77 32.0 19.1 57.5





#77

Bromobenzene

Concen: 21.074 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:156 Resp: 48024

Ion Ratio Lower Upper

156 100

77 164.9 88.4 265.3

158 97.0 48.0 144.0



#78

n-propylbenzene

Concen: 21.886 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.000 min

Lab File: VX037060.D




Acq: 17 Aug 2023 14:46


Tgt Ion: 91 Resp: 227389

Ion Ratio Lower Upper

91 100

120 23.0 11.5 34.4





#79

2-Chlorotoluene

Concen: 21.308 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 91 Resp: 141475

Ion Ratio Lower Upper

91 100

126 33.1 16.3 48.8



#80

1,3,5-Trimethylbenzene

Concen: 21.950 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX037060.D




Acq: 17 Aug 2023 14:46


Tgt Ion: 105 Resp: 170847

Ion Ratio Lower Upper

105 100

120 47.9 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 19.924 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:	75	Resp:	19502
Ion	Ratio	Lower	Upper
75	100		
53	99.7	90.5	135.7
89	47.7	36.9	55.3



#82
4-Chlorotoluene
Concen: 21.302 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion:	91	Resp:	163264
Ion	Ratio	Lower	Upper
91	100		
126	29.2	14.2	42.6





#83

tert-Butylbenzene

Concen: 21.494 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:119 Resp: 159549

Ion Ratio Lower Upper

119 100

91 61.1 31.2 93.6

134 23.9 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 21.449 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX037060.D




Acq: 17 Aug 2023 14:46


Tgt Ion:105 Resp: 168728

Ion Ratio Lower Upper

105 100

120 44.1 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 21.936 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion:105 Resp: 190191
 Ion Ratio Lower Upper
 105 100
 134 20.3 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 21.838 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46

Tgt Ion:119 Resp: 162994
 Ion Ratio Lower Upper
 119 100
 134 25.6 12.8 38.3
 91 25.3 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 20.711 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:146 Resp: 88362

Ion Ratio Lower Upper

146 100

111 42.0 22.1 66.1

148 62.9 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 20.758 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46




Tgt Ion:146 Resp: 89964


Ion Ratio Lower Upper

146 100

111 42.9 21.3 63.7

148 63.1 31.8 95.4





#89

n-Butylbenzene

Concen: 21.471 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX037060.D

Acq: 17 Aug 2023 14:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 91 Resp: 132866

Ion Ratio Lower Upper

91 100

92 53.5 27.4 82.0

134 26.1 12.6 37.8



#90

Hexachloroethane

Concen: 21.177 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX037060.D




Acq: 17 Aug 2023 14:46


Tgt Ion: 117 Resp: 26382

Ion Ratio Lower Upper

117 100

201 70.8 32.9 98.6









#91
 1,2-Dichlorobenzene
 Concen: 20.608 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion:146 Resp: 86096
 Ion Ratio Lower Upper
 146 100
 111 44.9 22.4 67.0
 148 63.7 31.6 95.0



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 20.234 ug/l
 RT: 12.945 min Scan# 1945
 Delta R.T. 0.000 min
 Lab File: VX037060.D
 Acq: 17 Aug 2023 14:46

Tgt Ion: 75 Resp: 14774
 Ion Ratio Lower Upper
 75 100
 155 84.2 36.1 108.5
 157 109.9 44.9 134.7









#93
1,2,4-Trichlorobenzene
Concen: 20.300 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:180 Resp: 49623
Ion Ratio Lower Upper
180 100
182 93.5 48.3 144.8
145 32.9 18.0 54.0



#94
Hexachlorobutadiene
Concen: 20.430 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46


Tgt Ion:225 Resp: 19293
Ion Ratio Lower Upper
225 100
223 63.3 31.1 93.2
227 62.0 32.5 97.5








#95
Naphthalene
Concen: 19.007 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:128 Resp: 203523

Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.3	15.5
129	11.0	8.8	13.2



#96
1,2,3-Trichlorobenzene
Concen: 20.377 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX037060.D
Acq: 17 Aug 2023 14:46

Tgt Ion:180 Resp: 49489

Ion	Ratio	Lower	Upper
180	100		
182	95.5	46.9	140.6
145	35.9	17.8	53.5

