

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081723\
 Data File : VX037064.D
 Acq On : 17 Aug 2023 16:22
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 18 06:08:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 05:51:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	122029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222673	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82989	47.173	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.340%
35) Dibromofluoromethane	5.385	113	71171	48.261	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.520%
50) Toluene-d8	8.647	98	267221	49.954	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.900%
62) 4-Bromofluorobenzene	11.079	95	104188	49.561	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.120%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.514	76	7987	2.500	ug/l #	93
94) Hexachlorobutadiene	13.725	225	6112	9.185	ug/l	98
95) Naphthalene	13.774	128	18671	2.475	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	4128	2.412	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081723\
Data File : VX037064.D
Acq On : 17 Aug 2023 16:22
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 18 06:08:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 05:51:31 2023
Response via : Initial Calibration

