

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081723\
 Data File : VX037065.D
 Acq On : 17 Aug 2023 17:41
 Operator : JC/MD
 Sample : BLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BLK

Quant Time: Aug 18 06:08:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 05:51:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	196556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	350139	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	332271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	152529	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	118368	41.771	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.540%
35) Dibromofluoromethane	5.385	113	106218	45.806	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.620%
50) Toluene-d8	8.647	98	420241	49.960	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.079	95	165696	50.126	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081723\
Data File : VX037065.D
Acq On : 17 Aug 2023 17:41
Operator : JC/MD
Sample : BLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BLK

Quant Time: Aug 18 06:08:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 05:51:31 2023
Response via : Initial Calibration

