

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
 Data File : VX023810.D
 Acq On : 18 Aug 2021 12:41
 Operator : JC/MD
 Sample : M3433-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-EB-20210812

Quant Time: Aug 18 13:44:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	130326	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	224928	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208736	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	94042	51.702	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.400%
35) Dibromofluoromethane	5.397	113	72500	48.699	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.400%
50) Toluene-d8	8.653	98	277663	49.003	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.000%
62) 4-Bromofluorobenzene	11.085	95	94278	45.198	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.400%
Target Compounds						
11) Tert butyl alcohol	2.971	59	20803	56.554	ug/l	97
16) Acetone	2.386	43	23444	28.826	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
Data File : VX023810.D
Acq On : 18 Aug 2021 12:41
Operator : JC/MD
Sample : M3433-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-EB-20210812

Quant Time: Aug 18 13:44:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

