

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
 Data File : VX023811.D
 Acq On : 18 Aug 2021 13:05
 Operator : JC/MD
 Sample : M3433-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-65-67

Quant Time: Aug 18 13:45:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	233370	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218746	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93156	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	97390	50.684	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.360%		
35) Dibromofluoromethane	5.397	113	78265	50.670	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.340%		
50) Toluene-d8	8.653	98	286198	48.682	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.360%		
62) 4-Bromofluorobenzene	11.085	95	104417	48.248	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.500%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	29235	34.028	ug/l	90
25) 2-Butanone	4.574	43	8191	6.294	ug/l	94
47) Bromodichloromethane	7.830	83	2295	1.074	ug/l	94
52) Toluene	8.726	92	1778	0.429	ug/l	93
60) Dibromochloromethane	9.525	129	1943	2.685	ug/l	87
84) 1,2,4-Trimethylbenzene	11.756	105	1636	0.281	ug/l	72
95) Naphthalene	13.780	128	22336	4.476	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
Data File : VX023811.D
Acq On : 18 Aug 2021 13:05
Operator : JC/MD
Sample : M3433-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-65-67

Quant Time: Aug 18 13:45:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

