

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
 Data File : VX023812.D
 Acq On : 18 Aug 2021 13:28
 Operator : JC/MD
 Sample : M3433-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-100-102

Quant Time: Aug 18 13:50:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131399	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	227575	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79291	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	95042	51.825	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.640%
35) Dibromofluoromethane	5.397	113	74365	49.371	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.740%
50) Toluene-d8	8.653	98	280079	48.855	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.700%
62) 4-Bromofluorobenzene	11.085	95	92385	43.775	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.560%

Target Compounds						Qvalue
16) Acetone	2.386	43	31691	38.648	ug/l #	88
25) 2-Butanone	4.574	43	7520	6.055	ug/l #	89
30) Chloroform	5.117	83	2736	1.003	ug/l	91
40) Benzene	6.056	78	4075	0.654	ug/l	100
52) Toluene	8.726	92	2260	0.560	ug/l	95
68) m/p-Xylenes	10.305	106	1263	0.431	ug/l	91
69) o-Xylene	10.647	106	856	0.298	ug/l	67
70) Styrene	10.665	104	994	0.212	ug/l	94
84) 1,2,4-Trimethylbenzene	11.750	105	2735	0.552	ug/l	94
95) Naphthalene	13.780	128	14606	3.736	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
Data File : VX023812.D
Acq On : 18 Aug 2021 13:28
Operator : JC/MD
Sample : M3433-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-100-102

Quant Time: Aug 18 13:50:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

