

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
 Data File : VX023815.D
 Acq On : 18 Aug 2021 14:39
 Operator : JC/MD
 Sample : PB138522TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138522TB

Quant Time: Aug 19 04:19:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132498	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	227221	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220157	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84325	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95806	51.808	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.620%	
35) Dibromofluoromethane	5.397	113	73446	48.836	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.680%	
50) Toluene-d8	8.653	98	289296	50.541	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.080%	
62) 4-Bromofluorobenzene	11.085	95	97892	46.457	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.920%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6010	7.269	ug/l	91
20) Methylene Chloride	2.794	84	9855	6.240	ug/l	93
43) Isopropyl Acetate	6.349	43	20599	5.091	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
Data File : VX023815.D
Acq On : 18 Aug 2021 14:39
Operator : JC/MD
Sample : PB138522TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138522TB

Quant Time: Aug 19 04:19:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

