

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
 Data File : VX023817.D
 Acq On : 18 Aug 2021 15:26
 Operator : JC/MD
 Sample : PB138522ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138522ZHE#02

Quant Time: Aug 19 04:19:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	137414	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	238214	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99493	51.877	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.760%	
35) Dibromofluoromethane	5.397	113	78621	49.865	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.720%	
50) Toluene-d8	8.653	98	302589	50.424	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.840%	
62) 4-Bromofluorobenzene	11.085	95	102841	46.553	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.100%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	7033	8.201	ug/l	# 82
20) Methylene Chloride	2.794	84	12367	7.550	ug/l	87
43) Isopropyl Acetate	6.354	43	21749	5.128	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
Data File : VX023817.D
Acq On : 18 Aug 2021 15:26
Operator : JC/MD
Sample : PB138522ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138522ZHE#02

Quant Time: Aug 19 04:19:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

