

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
 Data File : VX023819.D
 Acq On : 18 Aug 2021 16:14
 Operator : JC/MD
 Sample : PB138522ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138522ZHE#04

Quant Time: Aug 19 04:19:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	126454	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	223784	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	215259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87887	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	92975	52.680	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.360%
35) Dibromofluoromethane	5.397	113	72342	48.841	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.680%
50) Toluene-d8	8.653	98	280136	49.692	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.380%
62) 4-Bromofluorobenzene	11.085	95	96296	46.401	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.800%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6616	8.384	ug/l #	87
20) Methylene Chloride	2.794	84	12677	8.410	ug/l #	84
43) Isopropyl Acetate	6.354	43	20866	5.237	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
Data File : VX023819.D
Acq On : 18 Aug 2021 16:14
Operator : JC/MD
Sample : PB138522ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138522ZHE#04

Quant Time: Aug 19 04:19:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

