

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
 Data File : VX023820.D
 Acq On : 18 Aug 2021 16:37
 Operator : JC/MD
 Sample : PB138522ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138522ZHE#05

Quant Time: Aug 19 04:20:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127765	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	218327	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	91550	51.340	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.680%	
35) Dibromofluoromethane	5.397	113	71196	49.269	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.540%	
50) Toluene-d8	8.653	98	277403	50.437	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.880%	
62) 4-Bromofluorobenzene	11.085	95	93746	46.302	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.600%	
Target Compounds						Qvalue
16) Acetone	2.386	43	6769	8.490	ug/l	88
20) Methylene Chloride	2.794	84	13037	8.560	ug/l	88
43) Isopropyl Acetate	6.355	43	21752	5.595	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
Data File : VX023820.D
Acq On : 18 Aug 2021 16:37
Operator : JC/MD
Sample : PB138522ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138522ZHE#05

Quant Time: Aug 19 04:20:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

