

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081821\  
 Data File : VX023821.D  
 Acq On : 18 Aug 2021 17:01  
 Operator : JC/MD  
 Sample : PB138522ZHE#06  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB138522ZHE#06

Quant Time: Aug 19 04:20:21 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081721W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 17 11:57:51 2021  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc   | Units | Dev(Min) |
|-----------------------------|----------------|------|---------------------|--------|-------|----------|
| -----                       |                |      |                     |        |       |          |
| Internal Standards          |                |      |                     |        |       |          |
| 1) Pentafluorobenzene       | 5.562          | 168  | 127058              | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769          | 114  | 220642              | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.061         | 117  | 212594              | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024         | 152  | 80677               | 50.000 | ug/l  | 0.00     |
|                             |                |      |                     |        |       |          |
| System Monitoring Compounds |                |      |                     |        |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.970          | 65   | 92885               | 52.379 | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 78 - 117 |      | Recovery = 104.760% |        |       |          |
| 35) Dibromofluoromethane    | 5.397          | 113  | 72496               | 49.642 | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 75 - 124 |      | Recovery = 99.280%  |        |       |          |
| 50) Toluene-d8              | 8.653          | 98   | 278930              | 50.183 | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 92 - 112 |      | Recovery = 100.360% |        |       |          |
| 62) 4-Bromofluorobenzene    | 11.085         | 95   | 94362               | 46.117 | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 83 - 123 |      | Recovery = 92.240%  |        |       |          |
|                             |                |      |                     |        |       |          |
| Target Compounds            |                |      |                     |        |       | Qvalue   |
| 16) Acetone                 | 2.386          | 43   | 6357                | 8.017  | ug/l  | 91       |
| 20) Methylene Chloride      | 2.794          | 84   | 10602               | 7.000  | ug/l  | 89       |
| 43) Isopropyl Acetate       | 6.348          | 43   | 20998               | 5.345  | ug/l  | 93       |
| -----                       |                |      |                     |        |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081821\  
Data File : VX023821.D  
Acq On : 18 Aug 2021 17:01  
Operator : JC/MD  
Sample : PB138522ZHE#06  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
PB138522ZHE#06

Quant Time: Aug 19 04:20:21 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081721W.M  
Quant Title : SW846 8260  
QLast Update : Tue Aug 17 11:57:51 2021  
Response via : Initial Calibration

