

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
 Data File : VX023822.D
 Acq On : 18 Aug 2021 17:24
 Operator : JC/MD
 Sample : PB138522ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138522ZHE#07

Quant Time: Aug 19 04:20:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	231316	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92724	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97158	52.403	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.800%
35) Dibromofluoromethane	5.391	113	74579	48.712	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.420%
50) Toluene-d8	8.653	98	290243	49.809	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.620%
62) 4-Bromofluorobenzene	11.085	95	100946	47.058	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.120%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6426	7.752	ug/l	90
20) Methylene Chloride	2.794	84	9935	6.274	ug/l	90
43) Isopropyl Acetate	6.354	43	21437	5.205	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081821\
Data File : VX023822.D
Acq On : 18 Aug 2021 17:24
Operator : JC/MD
Sample : PB138522ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138522ZHE#07

Quant Time: Aug 19 04:20:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

