

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037076.D
 Acq On : 18 Aug 2023 11:55
 Operator : JC/MD
 Sample : 03980-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Aug 19 05:15:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	140997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	257409	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	250567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113074	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96716	47.580	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.160%
35) Dibromofluoromethane	5.391	113	81940	48.065	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.140%
50) Toluene-d8	8.647	98	308942	49.960	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.079	95	125587	51.679	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.360%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1178	1.481	ug/l #	79
20) Methylene Chloride	2.794	84	1150	0.647	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037076.D
Acq On : 18 Aug 2023 11:55
Operator : JC/MD
Sample : 03980-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Aug 19 05:15:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

