

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037079.D
 Acq On : 18 Aug 2023 13:07
 Operator : JC/MD
 Sample : 03980-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Aug 19 05:16:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	120985	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222446	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97956	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84394	48.385	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.780%
35) Dibromofluoromethane	5.391	113	70779	48.044	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.080%
50) Toluene-d8	8.646	98	263334	49.278	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.560%
62) 4-Bromofluorobenzene	11.079	95	107583	51.228	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.460%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	976	1.430	ug/l	98
20) Methylene Chloride	2.794	84	880	0.577	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037079.D
Acq On : 18 Aug 2023 13:07
Operator : JC/MD
Sample : 03980-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Aug 19 05:16:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

