

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037085.D
 Acq On : 18 Aug 2023 15:31
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 19 05:19:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125226	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	229508	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221154	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100955	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84517	46.815	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.620%
35) Dibromofluoromethane	5.391	113	71385	46.965	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.920%
50) Toluene-d8	8.647	98	275050	49.886	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.780%
62) 4-Bromofluorobenzene	11.079	95	110445	50.973	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.940%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	958	1.356	ug/l #	82
20) Methylene Chloride	2.788	84	1553	0.983	ug/l #	83
67) Ethyl Benzene	10.195	91	2475	0.301	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	1972	0.319	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037085.D
Acq On : 18 Aug 2023 15:31
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 19 05:19:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

