

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037086.D
 Acq On : 18 Aug 2023 15:55
 Operator : JC/MD
 Sample : 04024-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20230815

Quant Time: Aug 19 05:19:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136281	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	251647	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247838	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117239	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92627	47.145	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.280%		
35) Dibromofluoromethane	5.385	113	78649	47.191	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.380%		
50) Toluene-d8	8.647	98	302995	50.120	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	100.240%		
62) 4-Bromofluorobenzene	11.079	95	127162	53.525	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	107.040%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	1183	0.747	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.325	101	18578	12.896	ug/l	98
12) 1,1-Dichloroethene	2.318	96	9383	6.688	ug/l	81
16) Acetone	2.386	43	823	1.071	ug/l #	81
20) Methylene Chloride	2.788	84	1086	0.632	ug/l #	82
24) 1,1-Dichloroethane	3.611	63	2740	0.981	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	4081	2.171	ug/l	84
30) Chloroform	5.098	83	4745	1.513	ug/l	93
32) 1,1,1-Trichloroethane	5.385	97	1317	0.469	ug/l #	37
38) Carbon Tetrachloride	5.678	117	4162	1.618	ug/l	94
44) Trichloroethene	7.129	130	493063	256.061	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	1668	0.913	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037086.D
Acq On : 18 Aug 2023 15:55
Operator : JC/MD
Sample : 04024-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20230815

Quant Time: Aug 19 05:19:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

