

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037087.D
 Acq On : 18 Aug 2023 16:19
 Operator : JC/MD
 Sample : 04024-08DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20230815DL

Quant Time: Aug 19 05:19:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	118458	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	217312	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	212458	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101414	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82840	48.507	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	97.020%	
35) Dibromofluoromethane	5.391	113	69992	48.632	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.260%	
50) Toluene-d8	8.647	98	261767	50.142	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.280%	
62) 4-Bromofluorobenzene	11.079	95	109456	53.351	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	106.700%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	347	0.252	ug/l	82
9) 1,1,2-Trichlorotrifluo...	2.331	101	4805	3.837	ug/l	94
12) 1,1-Dichloroethene	2.319	96	2391	1.961	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	984	0.602	ug/l	74
38) Carbon Tetrachloride	5.690	117	1212	0.546	ug/l #	71
44) Trichloroethene	7.123	130	120056	72.199	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037087.D
Acq On : 18 Aug 2023 16:19
Operator : JC/MD
Sample : 04024-08DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20230815DL

Quant Time: Aug 19 05:19:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

