

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037088.D
 Acq On : 18 Aug 2023 16:43
 Operator : JC/MD
 Sample : 04024-02DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D2-20230815DL

Quant Time: Aug 19 05:20:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	127784	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238042	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87860	47.692	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.380%
35) Dibromofluoromethane	5.391	113	74958	47.547	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	8.647	98	283357	49.550	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.079	95	119310	53.090	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.180%
Target Compounds						
27) cis-1,2-Dichloroethene	4.489	96	4472	2.537	ug/l	78
44) Trichloroethene	7.129	130	51062	28.033	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037088.D
Acq On : 18 Aug 2023 16:43
Operator : JC/MD
Sample : 04024-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D2-20230815DL

Quant Time: Aug 19 05:20:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

