

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037089.D
 Acq On : 18 Aug 2023 17:07
 Operator : JC/MD
 Sample : 04024-15DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20230815DL

Quant Time: Aug 19 05:20:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	112393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205433	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92538	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79278	48.927	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.860%
35) Dibromofluoromethane	5.385	113	65916	48.449	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.900%
50) Toluene-d8	8.647	98	247433	50.137	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.280%
62) 4-Bromofluorobenzene	11.079	95	100742	51.943	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.880%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	2480	2.087	ug/l	98
27) cis-1,2-Dichloroethene	4.507	96	865	0.558	ug/l	85
44) Trichloroethene	7.129	130	224197	142.624	ug/l	98
64) Tetrachloroethene	9.275	164	1779	1.206	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037089.D
Acq On : 18 Aug 2023 17:07
Operator : JC/MD
Sample : 04024-15DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20230815DL

Quant Time: Aug 19 05:20:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

