

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037090.D
 Acq On : 18 Aug 2023 17:32
 Operator : JC/MD
 Sample : 04024-19DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW17811-20230815DL

Quant Time: Aug 19 05:20:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	116315	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	211955	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	203979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95119	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80920	48.256	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.520%
35) Dibromofluoromethane	5.391	113	67352	47.981	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.960%
50) Toluene-d8	8.647	98	252597	49.608	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.220%
62) 4-Bromofluorobenzene	11.079	95	102855	51.401	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.800%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	389	0.325	ug/l #	71
27) cis-1,2-Dichloroethene	4.489	96	14088	8.780	ug/l	83
44) Trichloroethene	7.129	130	179986	110.975	ug/l	98
45) 1,2-Dichloropropane	7.434	63	9408	6.671	ug/l	98
64) Tetrachloroethene	9.275	164	9968	6.699	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037090.D
Acq On : 18 Aug 2023 17:32
Operator : JC/MD
Sample : 04024-19DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW17811-20230815DL

Quant Time: Aug 19 05:20:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

