

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037091.D
 Acq On : 18 Aug 2023 17:56
 Operator : JC/MD
 Sample : 04024-11DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20230815DL

Quant Time: Aug 19 05:21:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	110011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203994	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95089	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76968	48.529	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.060%
35) Dibromofluoromethane	5.391	113	64519	47.756	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.520%
50) Toluene-d8	8.647	98	244873	49.968	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.940%
62) 4-Bromofluorobenzene	11.079	95	101925	52.924	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.840%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	3294	2.833	ug/l	96
12) 1,1-Dichloroethene	2.319	96	516	0.456	ug/l #	81
27) cis-1,2-Dichloroethene	4.501	96	630	0.415	ug/l	88
30) Chloroform	5.093	83	592	0.234	ug/l #	61
44) Trichloroethene	7.129	130	65517	41.973	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037091.D
Acq On : 18 Aug 2023 17:56
Operator : JC/MD
Sample : 04024-11DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20230815DL

Quant Time: Aug 19 05:21:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

