

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037097.D
 Acq On : 18 Aug 2023 20:19
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 21 01:34:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	123067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221672	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214918	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94514	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84067	47.382	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.760%
35) Dibromofluoromethane	5.385	113	70177	47.802	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.600%
50) Toluene-d8	8.647	98	267589	50.249	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.500%
62) 4-Bromofluorobenzene	11.079	95	107312	51.278	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.560%
Target Compounds						
94) Hexachlorobutadiene	13.725	225	2140	3.078	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037097.D
Acq On : 18 Aug 2023 20:19
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 21 01:34:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

