

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037100.D
 Acq On : 19 Aug 2023 01:40
 Operator : JC/MD
 Sample : VX0818WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0818WBL02

Quant Time: Aug 21 01:36:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	158228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	278812	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	114850	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	122621	53.754	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.500%
35) Dibromofluoromethane	5.385	113	95639	51.795	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.580%
50) Toluene-d8	8.647	98	348204	51.986	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.980%
62) 4-Bromofluorobenzene	11.079	95	130863	49.716	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037100.D
Acq On : 19 Aug 2023 01:40
Operator : JC/MD
Sample : VX0818WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0818WBL02

Quant Time: Aug 21 01:36:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

