

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037107.D
 Acq On : 19 Aug 2023 04:26
 Operator : JC/MD
 Sample : 04047-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20230816

Quant Time: Aug 21 01:40:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	165230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	281475	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	124721	52.358	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.720%	
35) Dibromofluoromethane	5.385	113	97467	52.285	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.580%	
50) Toluene-d8	8.646	98	361913	53.522	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.040%	
62) 4-Bromofluorobenzene	11.079	95	136801	51.480	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.960%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037107.D
Acq On : 19 Aug 2023 04:26
Operator : JC/MD
Sample : 04047-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20230816

Quant Time: Aug 21 01:40:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

