

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037109.D
 Acq On : 19 Aug 2023 05:13
 Operator : JC/MD
 Sample : 04024-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20230815

Quant Time: Aug 21 01:41:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	158992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	277854	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255664	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119212	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	119286	52.041	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.080%		
35) Dibromofluoromethane	5.385	113	93266	50.684	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.360%		
50) Toluene-d8	8.647	98	341455	51.155	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.300%		
62) 4-Bromofluorobenzene	11.079	95	130561	49.772	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 99.540%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	13086	7.786	ug/l	96
12) 1,1-Dichloroethene	2.319	96	1613	0.986	ug/l	81
27) cis-1,2-Dichloroethene	4.495	96	2389	1.089	ug/l	85
38) Carbon Tetrachloride	5.684	117	968	0.341	ug/l #	81
44) Trichloroethene	7.123	130	1036122	487.333	ug/l	99
64) Tetrachloroethene	9.275	164	9724	5.214	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037109.D
Acq On : 19 Aug 2023 05:13
Operator : JC/MD
Sample : 04024-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20230815

Quant Time: Aug 21 01:41:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

