

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081823\  
 Data File : VX037114.D  
 Acq On : 19 Aug 2023 07:11  
 Operator : JC/MD  
 Sample : 04025-04  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 49 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE126D3-20230814

Quant Time: Aug 21 01:43:31 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081723W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 18 09:30:38 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 136069   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.757  | 114   | 246340   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 225773   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.018 | 152   | 106364   | 50.000   | ug/l       | 0.00     |
|                             |        |       |          |          |            |          |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 108939   | 55.534   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 111.060% |          |
| 35) Dibromofluoromethane    | 5.385  | 113   | 84095    | 51.546   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 103.100% |          |
| 50) Toluene-d8              | 8.647  | 98    | 304029   | 51.375   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 102.740% |          |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 117160   | 50.377   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 100.760% |          |
|                             |        |       |          |          |            |          |
| Target Compounds            |        |       |          |          |            | Qvalue   |
| 44) Trichloroethene         | 7.123  | 130   | 3711     | 1.969    | ug/l       | 89       |
| 64) Tetrachloroethene       | 9.275  | 164   | 1919     | 1.165    | ug/l #     | 80       |
| -----                       |        |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081823\  
Data File : VX037114.D  
Acq On : 19 Aug 2023 07:11  
Operator : JC/MD  
Sample : 04025-04  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 49 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE126D3-20230814

Quant Time: Aug 21 01:43:31 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081723W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 18 09:30:38 2023  
Response via : Initial Calibration

