

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037116.D
 Acq On : 19 Aug 2023 07:58
 Operator : JC/MD
 Sample : 04025-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D1-20230814

Quant Time: Aug 21 01:44:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	159801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	279307	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112006	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	122379	53.120	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.240%	
35) Dibromofluoromethane	5.385	113	96291	52.055	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.120%	
50) Toluene-d8	8.647	98	349424	52.076	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.160%	
62) 4-Bromofluorobenzene	11.079	95	130425	49.462	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.920%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	2895	1.714	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	2495	1.132	ug/l	82
44) Trichloroethene	7.123	130	492012	230.211	ug/l	99
64) Tetrachloroethene	9.275	164	3192	1.778	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
Data File : VX037116.D
Acq On : 19 Aug 2023 07:58
Operator : JC/MD
Sample : 04025-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D1-20230814

Quant Time: Aug 21 01:44:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

