



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081823\
 Data File : VX037069.D
 Acq On : 18 Aug 2023 08:19
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Title : SW846 8260
 Last Update : Fri Aug 18 09:30:38 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	10087	PASS
75	95	30	60	51.4	29925	PASS
95	95	100	100	100.0	58200	PASS
96	95	5	9	6.7	3875	PASS
173	174	0.00	2	0.9	368	PASS
174	95	50	100	73.3	42640	PASS
175	174	5	9	7.5	3188	PASS
176	174	95	101	97.1	41411	PASS
177	176	5	9	6.8	2807	PASS