

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081825\
 Data File : VX047382.D
 Acq On : 18 Aug 2025 15:14
 Operator : JC/MD
 Sample : Q2815-06RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-10P-NRE

Quant Time: Aug 19 01:43:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	232145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	447155	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	455584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	236692	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	177252	54.557	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 109.120%		
35) Dibromofluoromethane	5.397	113	146545	50.497	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.000%		
50) Toluene-d8	8.653	98	523865	50.504	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.000%		
62) 4-Bromofluorobenzene	11.079	95	223111	58.288	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 116.580%		
Target Compounds						
16) Acetone	2.392	43	8786	7.194	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081825\
Data File : VX047382.D
Acq On : 18 Aug 2025 15:14
Operator : JC/MD
Sample : Q2815-06RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-10P-NRE

Quant Time: Aug 19 01:43:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

