

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081825\
 Data File : VX047397.D
 Acq On : 18 Aug 2025 20:35
 Operator : JC/MD
 Sample : Q2888-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-210

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/19/2025
 Supervised By : Mahesh Dadoda 08/20/2025

Quant Time: Aug 19 01:50:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	275192	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	542301	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	527857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	264358	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	209777	54.468	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 108.940%			
35) Dibromofluoromethane	5.404	113	175898	49.977	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.960%			
50) Toluene-d8	8.653	98	641640	51.005	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.020%			
62) 4-Bromofluorobenzene	11.079	95	257810	55.536	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 111.080%			
Target Compounds						
16) Acetone	2.392	43	65972	45.571	ug/l	98
18) Methyl Acetate	2.721	43	28218	7.522	ug/l #	90
20) Methylene Chloride	2.806	84	21956	5.727	ug/l	99
37) Ethyl Acetate	4.763	43	12072m	2.093	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081825\
Data File : VX047397.D
Acq On : 18 Aug 2025 20:35
Operator : JC/MD
Sample : Q2888-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-210

Manual Integrations
APPROVED

Reviewed By : John Carlone 08/19/2025
Supervised By : Mahesh Dadoda 08/20/2025

Quant Time: Aug 19 01:50:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

