

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081919\
 Data File : VX011635.D
 Acq On : 19 Aug 2019 16:29
 Operator : JC/SP
 Sample : K4368-04DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PH2-0133DL

Manual Integrations
 APPROVED

MMDadoda
 8/20/2019 3:41:31 PM

Quant Time: Aug 21 09:22:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	443274	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	608405	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	535356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	292652	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	208573	50.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.48%	
35) Dibromofluoromethane	5.49	113	179313	45.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.12%	
50) Toluene-d8	8.71	98	691896	55.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.60%	
62) 4-Bromofluorobenzene	11.13	95	238748	45.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.24%	
Target Compounds						
39) Methylcyclohexane	7.46	83	10741	1.672	ug/l	98
67) Ethyl Benzene	10.25	91	439156	23.045	ug/l	99
68) m/p-Xylenes	10.35	106	753943	103.069	ug/l	95
69) o-Xylene	10.69	106	5113	0.714	ug/l	100
73) Isopropylbenzene	11.02	105	58861	3.100	ug/l	98
78) n-propylbenzene	11.36	91	82601	4.034	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	160541	10.064	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	583691	36.447	ug/l	99
85) sec-Butylbenzene	11.94	105	8464m	0.462	ug/l	
95) Naphthalene	13.83	128	224940	10.954	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081919\
Data File : VX011635.D
Acq On : 19 Aug 2019 16:29
Operator : JC/SP
Sample : K4368-04DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PH2-0133DL

Manual Integrations
APPROVED

MMDadoda
8/20/2019 3:41:31 PM

Quant Time: Aug 21 09:22:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

